

Telestroke BETA Test Schedule

Action	Success Criteria	Result
Login into the Laptop using the credentials provided	Login achieved without error messages etc	
Live Video, Burnbank, iPanel and NHS email Icons are present on desktop	All 4 Icons are present on Desktop	
Quick user guide and full user guide are accessible from the desktop	Both Guides can be opened and viewed	
Using the Icon on the Desktop login to NHS email, send an email with the subject heading "test" to XXX@Imerja.com	Application is accessible and email is sent	
Confirm access to recordings by selecting Videos on ipanel button.	Test video is present and can be viewed	
Using iPanel favourites setup a call to Blackpool Victoria Hospital - Confirm acceptable Image and Camera Control	See VC Success Criteria	
Using iPanel favourites setup a call to Cumberland Infirmary - Confirm acceptable Image and Camera Control	See VC Success Criteria	
Using iPanel favourites setup a call to Furness General Hospital - Confirm acceptable Image and Camera Control	See VC Success Criteria	
Using iPanel favourites setup a call to Royal Blackburn Hospital - Confirm acceptable Image and Camera Control	See VC Success Criteria	
Using iPanel favourites setup a call to Royal Lancaster Hospital - Confirm acceptable Image and Camera Control	See VC Success Criteria	
Using iPanel favourites setup a call to Royal Preston Hospital - Confirm acceptable Image and Camera Control	See VC Success Criteria	
Using iPanel favourites setup a call to Southport & Formby Hospital - Confirm acceptable Image and Camera Control	See VC Success Criteria	
Using iPanel favourites setup a call to West Cumberland Hospital - Confirm acceptable Image and Camera Control	See VC Success Criteria	
Using the Icon on the Desktop login to Burnbank and access PACs Image	Application is accessible and Images can be viewed	
Check NHS email inbox for a replay to the test message	Test message present	
Call the Contact Centre and confirm completion of test.	Contact Centre answer call and log notification	

VC Success Criteria - The A&E Carts will not be attended too during the test, however they should be contactable and positioned facing an object that will provide an image for the Clinician. It may not be possible to access all Carts and the vista available to the camera may vary. However where possible (and at least too two carts) the Clinician should confirm that a stable and clear image is visible on the laptop. Via control buttons on the CMA desktop it should be possible to pan the Camera on the Cart approximately 180 degrees from left to right, 45 degrees up and down and zoom (x12) the image. Control should be such that a Clinician should be able to centre and size an image on the laptop screen.